

Deutsche Akkreditierungsstelle

Annex to the Accreditation Certificate D-K-19592-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 11.02.2025

Date of issue: 11.06.2025

Holder of accreditation certificate:

Amphenol Advanced Sensors Germany GmbH Sinsheimer Straße 6, 75179 Pforzheim

with the location

Amphenol Advanced Sensors Germany GmbH Sinsheimer Straße 6, 75179 Pforzheim

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and they conform to the principles of DIN EN ISO 9001.

Calibration in the fields:

Thermodynamic quantities

Humidity quantities

Devices for relative humidity

Temperature quantities

- Direct reading thermometers
- Temperature transmitters, data loggers
- Resistance thermometers

This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at https://www.dakks.de.

Abbreviations used: see last page



Annex to the Accreditation Certificate D-K-19592-01-00

The calibration laboratory is permitted, without being required to inform and obtain prior approval from DAkkS, to use calibration standards or equivalent calibration procedures listed here with different issue dates.

The calibration laboratory maintains a current list of all calibration standards / equivalent calibration procedures within the flexible scope of accreditation.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range		Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks	
Humidity quantities Data logger for relative humidity	15 % to	30 %	DKD-R 5-8:2019	0.3 %	Comparison with dew point mirror Measurement uncertainty expressed in relative humidity	
	> 30 % to	70 %	in 2-pressure-generator temperature range: 10 °C to 60 °C	0.6 %		
	> 70 % to	95 %		0.8 %		
Resistance thermometers; direct reading thermometers and temperature transmitters with resistance sensor	–196°	С	DKD-R 5-1:2018 in copper block in liquid nitrogen	10 mK	Comparison with standard resistance thermometer (SPRT)	
	–65 ℃ to	–55 °C	DKD-R 5-1:2018 in alcohol bath	8 mK	interpolation of the characteristic curve -according to DKD-R 5-6:2018	
	−0,1 °C to	0,1 °C	DKD-R 5-1:2018 in water bath	5 mK		
	95 °C to	105 °C	DKD-R 5-1:2018 in oil bath	7 mK		
	255 °C to	265 °C	DKD-R 5-1:2018 in salt bath	9 mK		
	415 °C to	425 °C		12 mK		
	−85 °C to	<-65 °C	DKD-R 5-1:2018 in liquid baths	10 mK		
	>-55 °C to	<-0,1 °C		10 mK		
	> 0,1 °C to	< 95 °C		10 mK		
	> 105 °C to	< 255 °C		10 mK		
	> 265 °C to	< 415 °C		15 mK		
Temperature data loggers with resistance sensor	−85 °C to	<-60 °C	DKD-R 5-1:2018 in liquid baths	50 mK	Comparison with reference resistance thermometer (IPRT) interpolation of the characteristic curve according to DKD-R 5-6:2018	
	−60 °C to	260 °C		25 mK		
	> 260 °C to	400 °C		50 mK		

Abbreviations used:

CMC Calibration and measurement capabilities

DIN Deutsches Institut für Normung e.V. – German institute for standardization

DKD-R Guideline of Deutscher Kalibrierdienst (DKD), published by Physikalisch-Technische Bundesanstalt

EN Europäische Norm – European Standard
 IEC International Electrotechnical Commission
 ISO International Organization for Standardisation

Valid from: 11.02.2025 Date of issue: 11.06.2025